

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
 Data File : VU042779.D
 Acq On : 25 Mar 2021 12:54
 Operator : SY/MD
 Sample : PB1355335ZHE#01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB1355335ZHE#01

Quant Time: Mar 26 02:05:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	112200	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	180360	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.426	117	178374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	84808	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	87941	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
35) Dibromofluoromethane	5.298	113	61731	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
50) Toluene-d8	7.906	98	228028	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
62) 4-Bromofluorobenzene	10.642	95	83414	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
Target Compounds						Qvalue
16) Acetone	2.639	43	19882	8.94	ug/l	99
20) Methylene Chloride	3.054	84	6603	5.13	ug/l	97
43) Isopropyl Acetate	5.922	43	13407	3.22	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
Data File : VU042779.D
Acq On : 25 Mar 2021 12:54
Operator : SY/MD
Sample : PB1355335ZHE#01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB1355335ZHE#01

Quant Time: Mar 26 02:05:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

