

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
 Data File : VU042785.D
 Acq On : 25 Mar 2021 15:13
 Operator : SY/MD
 Sample : M1791-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXTERIOR-SLAB-BRICK-2

Quant Time: Mar 26 02:06:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	108570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	171298	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	170744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	81902	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	86715	50.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.20%	
35) Dibromofluoromethane	5.298	113	61179	47.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.72%	
50) Toluene-d8	7.906	98	218109	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
62) 4-Bromofluorobenzene	10.639	95	78973	45.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.38%	
Target Compounds						
16) Acetone	2.639	43	28635	20.15	ug/l	96
18) Methyl Acetate	2.961	43	2255	1.14	ug/l #	78
20) Methylene Chloride	3.057	84	4428	3.55	ug/l	97
43) Isopropyl Acetate	5.919	43	5998	1.51	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
Data File : VU042785.D
Acq On : 25 Mar 2021 15:13
Operator : SY/MD
Sample : M1791-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXTERIOR-SLAB-BRICK-2

Quant Time: Mar 26 02:06:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

