

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
 Data File : VU042787.D
 Acq On : 25 Mar 2021 15:59
 Operator : SY/MD
 Sample : M1776-01 10X
 Misc : 1.0g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 110

Quant Time: Mar 26 02:06:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	109233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	168397	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	171954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	117656	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	83356	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	5.298	113	59611	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	7.906	98	216598	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	10.642	95	101112	58.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.70%	
Target Compounds					Qvalue	
39) Methylcyclohexane	6.767	83	3164	1.66	ug/l	97
52) Toluene	7.976	92	66646	21.88	ug/l	98
64) Tetrachloroethene	8.558	164	17177	12.37	ug/l	97
67) Ethyl Benzene	9.578	91	539296	86.00	ug/l	98
68) m/p-Xylenes	9.703	106	1172397	507.28	ug/l	98
69) o-Xylene	10.108	106	359596	160.75	ug/l	97
73) Isopropylbenzene	10.491	105	29588	3.97	ug/l	99
78) n-propylbenzene	10.912	91	111625	12.73	ug/l	99
80) 1,3,5-Trimethylbenzene	11.095	105	170302	26.36	ug/l	100
84) 1,2,4-Trimethylbenzene	11.475	105	698090	106.99	ug/l	99
85) sec-Butylbenzene	11.651	105	88814	12.14	ug/l	99
86) p-Isopropyltoluene	11.803	119	59163	8.84	ug/l	99
89) n-Butylbenzene	12.214	91	313239	49.58	ug/l #	52
95) Naphthalene	14.095	128	147223	20.08	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
Data File : VU042787.D
Acq On : 25 Mar 2021 15:59
Operator : SY/MD
Sample : M1776-01 10X
Misc : 1.0g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
110

Quant Time: Mar 26 02:06:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

