

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032522\
 Data File : VU047676.D
 Acq On : 25 Mar 2022 16:48
 Operator : SY/MD
 Sample : VU0325WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK072

Quant Time: Mar 26 02:50:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 26 02:50:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	283063	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	292543	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	159245	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	99320	48.386	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.780%
7) Chloroethane-d5	1.915	69	88246	54.522	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.040%
11) 1,1-Dichloroethene-d2	2.571	63	140536	37.421	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.840%
21) 2-Butanone-d5	4.639	46	211229	119.890	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.890%
24) Chloroform-d	5.066	84	193257	51.455	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.900%
26) 1,2-Dichloroethane-d4	5.706	65	126885	55.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.160%
32) Benzene-d6	5.732	84	398840	54.181	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.360%
36) 1,2-Dichloropropane-d6	6.693	67	126920	56.142	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.280%
41) Toluene-d8	7.899	98	368746	51.415	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.820%
43) trans-1,3-Dichloroprop...	8.182	79	59430	51.890	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.780%
47) 2-Hexanone-d5	8.635	63	154247	116.553	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	116.550%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	204311	52.661	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.320%
66) 1,2-Dichlorobenzene-d4	12.195	152	162408	57.325	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032522\
Data File : VU047676.D
Acq On : 25 Mar 2022 16:48
Operator : SY/MD
Sample : VU0325WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK072

Quant Time: Mar 26 02:50:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 26 02:50:24 2022
Response via : Initial Calibration

