

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032524\
 Data File : VU058218.D
 Acq On : 25 Mar 2024 18:36
 Operator : MD/SY
 Sample : P1796-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 26 01:43:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 26 01:41:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	67871	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	67225	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	29162	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	23968	4.179	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.600%	
7) Chloroethane-d5	1.907	69	23509	5.022	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	100.400%	
11) 1,1-Dichloroethene-d2	2.560	65	10653	4.474	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	89.400%	
20) 2-Butanone-d5	4.628	46	53858	52.856	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.720%	
24) Chloroform-d	5.052	84	48998	4.778	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.600%	
26) 1,2-Dichloroethane-d4	5.692	65	24600	5.131	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.600%	
32) Benzene-d6	5.718	84	99824	4.572	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.400%	
36) 1,2-Dichloropropane-d6	6.682	67	31394	4.878	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.600%	
41) Toluene-d8	7.891	98	82322	4.256	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.200%	
43) trans-1,3-Dichloroprop...	8.174	79	9159	4.442	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.800%	
46) 2-Hexanone-d5	8.627	63	39768	47.624	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.240%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	20467	4.882	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	26653	4.908	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.200%	
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	2652	0.550	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032524\
Data File : VU058218.D
Acq On : 25 Mar 2024 18:36
Operator : MD/SY
Sample : P1796-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 26 01:43:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 26 01:41:16 2024
Response via : Initial Calibration

