

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032524\
 Data File : VU058211.D
 Acq On : 25 Mar 2024 15:49
 Operator : MD/SY
 Sample : VU0325WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK154

Quant Time: Mar 26 01:41:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 26 01:41:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	71372	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	68751	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	30363	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	24375	4.041	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.907	69	22435	4.557	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.200%
11) 1,1-Dichloroethene-d2	2.560	65	9936	3.968	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.400%
20) 2-Butanone-d5	4.624	46	52458	48.957	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.920%
24) Chloroform-d	5.052	84	48423	4.491	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.692	65	23604	4.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.718	84	98520	4.412	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.679	67	30884	4.693	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.891	98	83120	4.202	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	8.174	79	9494	4.502	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.627	63	36955	43.273	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.540%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20138	4.697	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	26522	4.691	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032524\
Data File : VU058211.D
Acq On : 25 Mar 2024 15:49
Operator : MD/SY
Sample : VU0325WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK154

Quant Time: Mar 26 01:41:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 26 01:41:16 2024
Response via : Initial Calibration

