

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032524\  
 Data File : VU058213.D  
 Acq On : 25 Mar 2024 16:39  
 Operator : MD/SY  
 Sample : P1841-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 C0AK9

Quant Time: Mar 26 01:42:24 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR032124WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Mar 26 01:41:16 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 6.242  | 114            | 73754    | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.412  | 117            | 70255    | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.807 | 152            | 30828    | 5.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.595  | 65             | 25536    | 4.097  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 82.000%  |          |
| 7) Chloroethane-d5            | 1.907  | 69             | 24219    | 4.761  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 95.200%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.563  | 65             | 10933    | 4.225  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 84.400%  |          |
| 20) 2-Butanone-d5             | 4.624  | 46             | 56780    | 51.279 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 102.560% |          |
| 24) Chloroform-d              | 5.052  | 84             | 51387    | 4.612  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 92.200%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.695  | 65             | 25609    | 4.916  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 98.400%  |          |
| 32) Benzene-d6                | 5.717  | 84             | 101722   | 4.458  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 89.200%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.682  | 67             | 32122    | 4.776  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 95.600%  |          |
| 41) Toluene-d8                | 7.891  | 98             | 87377    | 4.322  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 86.400%  |          |
| 43) trans-1,3-Dichloroprop... | 8.174  | 79             | 9696     | 4.500  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 90.000%  |          |
| 46) 2-Hexanone-d5             | 8.627  | 63             | 40830    | 46.787 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 93.580%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.749 | 84             | 21390    | 4.882  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 97.600%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.187 | 152            | 28531    | 4.970  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 99.400%  |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032524\  
Data File : VU058213.D  
Acq On : 25 Mar 2024 16:39  
Operator : MD/SY  
Sample : P1841-04  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0AK9

Quant Time: Mar 26 01:42:24 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR032124WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Mar 26 01:41:16 2024  
Response via : Initial Calibration

