

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032719\
 Data File : VU030473.D
 Acq On : 27 Mar 2019 11:03
 Operator : JC/SP
 Sample : VU0327WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK59

Quant Time: Mar 28 01:38:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 28 01:25:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	385726	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	372658	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	148141	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	160444	51.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.38%
7) Chloroethane-d5	1.68	69	132721	53.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.88%
11) 1,1-Dichloroethene-d2	2.27	63	226812	38.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.04%
21) 2-Butanone-d5	4.17	46	302578	103.69	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.69%
24) Chloroform-d	4.64	84	275379	51.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.98%
26) 1,2-Dichloroethane-d4	5.31	65	184128	52.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.22%
32) Benzene-d6	5.34	84	545345	51.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.06%
36) 1,2-Dichloropropane-d6	6.32	67	198325	54.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.48%
41) Toluene-d8	7.56	98	472029	49.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.36%
43) trans-1,3-Dichloropropene-	7.85	79	84546	49.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.34%
47) 2-Hexanone-d5	8.31	63	196442	99.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.14%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	265929	50.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.66%
64) 1,2-Dichlorobenzene-d4	11.86	152	158280	53.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.64%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU032719\
Data File : VU030473.D
Acq On : 27 Mar 2019 11:03
Operator : JC/SP
Sample : VU0327WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK59

Quant Time: Mar 28 01:38:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 01:25:46 2019
Response via : Initial Calibration

