

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032824\
 Data File : VU058235.D
 Acq On : 28 Mar 2024 13:22
 Operator : MD/SY
 Sample : VU0328WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK156

Quant Time: Mar 29 01:09:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 29 01:09:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	64112	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61894	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	26083	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	22988	4.243	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.800%
7) Chloroethane-d5	1.908	69	21751	4.918	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.563	65	9903	4.402	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.000%
20) 2-Butanone-d5	4.628	46	54206	56.317	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.640%
24) Chloroform-d	5.055	84	47508	4.905	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.695	65	23184	5.120	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.400%
32) Benzene-d6	5.721	84	92046	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.682	67	30127	5.085	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.600%
41) Toluene-d8	7.891	98	74231	4.168	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
43) trans-1,3-Dichloroprop...	8.174	79	8116	4.275	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.628	63	38912	50.613	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.220%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20144	5.219	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	23785	4.897	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032824\
Data File : VU058235.D
Acq On : 28 Mar 2024 13:22
Operator : MD/SY
Sample : VU0328WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK156

Quant Time: Mar 29 01:09:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 29 01:09:13 2024
Response via : Initial Calibration

