

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032921\
 Data File : VU042818.D
 Acq On : 29 Mar 2021 16:02
 Operator : SY/MD
 Sample : M1846-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAR23

Quant Time: Mar 30 01:16:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 30 01:12:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	91899	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	92664	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	44445	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	22143	3.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.20%
7) Chloroethane-d5	1.919	69	20355	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.80%
11) 1,1-Dichloroethene-d2	2.572	63	38622	2.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.80%#
20) 2-Butanone-d5	4.649	46	95827	47.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.28%
24) Chloroform-d	5.073	84	51038	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.20%
26) 1,2-Dichloroethane-d4	5.710	65	33085	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
32) Benzene-d6	5.735	84	91476	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.60%
36) 1,2-Dichloropropane-d6	6.700	67	30792	4.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.00%
41) Toluene-d8	7.906	98	82147	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.00%
43) trans-1,3-Dichloroprop...	8.189	79	12657	3.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.00%
46) 2-Hexanone-d5	8.649	63	67381	41.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.20%
57) 1,1,2,2-Tetrachloroeth...	10.767	84	26596	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	12.201	152	35375	4.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.80%
Target Compounds						Qvalue
13) Acetone	2.658	43	19027	12.27	ug/L	96
25) Chloroform	5.099	83	3033	0.22	ug/L	88
34) Trichloroethene	6.552	95	36842	4.58	ug/L	94
47) Tetrachloroethene	8.562	164	699	0.11	ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032921\
Data File : VU042818.D
Acq On : 29 Mar 2021 16:02
Operator : SY/MD
Sample : M1846-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAR23

Quant Time: Mar 30 01:16:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 30 01:12:10 2021
Response via : Initial Calibration









#13
Acetone
Concen: 12.27 ug/L
RT: 2.658 min Scan# 410
Delta R.T. 0.000 min
Lab File: VU042818.D
Acq: 29 Mar 2021 16:02




Instrument :
MSVOA_U
ClientSampleId :
YAR23


Tgt Ion: 43 Resp: 19027
Ion Ratio Lower Upper
43 100
58 30.1 0.0 55.8



#25
Chloroform
Concen: 0.22 ug/L
RT: 5.099 min Scan# 1169
Delta R.T. 0.000 min
Lab File: VU042818.D
Acq: 29 Mar 2021 16:02

Tgt Ion: 83 Resp: 3033
Ion Ratio Lower Upper
83 100
85 55.8 45.7 84.9





#34

Trichloroethene

Concen: 4.58 ug/L

RT: 6.552 min Scan# 1621

Delta R.T. 0.000 min

Lab File: VU042818.D

Acq: 29 Mar 2021 16:02

Tgt Ion: 95 Resp: 36842





Ion Ratio Lower Upper

95 100

97 60.1 44.4 82.4

132 90.1 68.0 126.4

130 93.1 69.2 128.4



#47

Tetrachloroethene

Concen: 0.11 ug/L

RT: 8.562 min Scan# 2246

Delta R.T. 0.000 min

Lab File: VU042818.D

Acq: 29 Mar 2021 16:02

Tgt Ion: 164 Resp: 699

Ion Ratio Lower Upper

164 100

129 78.3 64.3 119.5

131 100.0 62.2 115.6

166 121.5 88.6 164.6

