

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
 Data File : VU058261.D
 Acq On : 29 Mar 2024 11:50
 Operator : MD/SY
 Sample : VU0329WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK157

Quant Time: Mar 29 23:41:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 29 23:40:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	71906	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	69114	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	27795	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21594	3.554	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.000%
7) Chloroethane-d5	1.911	69	21788	4.393	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.800%
11) 1,1-Dichloroethene-d2	2.563	65	9072	3.596	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.000%
20) 2-Butanone-d5	4.634	46	57234	53.017	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.040%
24) Chloroform-d	5.055	84	46281	4.260	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	5.695	65	23169	4.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	5.721	84	90307	4.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	6.682	67	29672	4.485	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.600%
41) Toluene-d8	7.891	98	70829	3.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.200%
43) trans-1,3-Dichloroprop...	8.177	79	8040	3.793	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	8.631	63	39596	46.122	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.240%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	19742	4.580	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	23149	4.472	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
Data File : VU058261.D
Acq On : 29 Mar 2024 11:50
Operator : MD/SY
Sample : VU0329WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK157

Quant Time: Mar 29 23:41:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 29 23:40:26 2024
Response via : Initial Calibration

