

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
 Data File : VU058265.D
 Acq On : 29 Mar 2024 13:33
 Operator : MD/SY
 Sample : P1902-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOC4

Quant Time: Mar 29 23:42:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 29 23:40:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	75368	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	74446	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	31559	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	30275	4.753	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.000%	
7) Chloroethane-d5	1.907	69	27396	5.270	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	105.400%	
11) 1,1-Dichloroethene-d2	2.563	65	12736	4.816	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	96.400%	
20) 2-Butanone-d5	4.618	46	47825	42.267	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	84.540%	
24) Chloroform-d	5.052	84	52855	4.642	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	5.695	65	24762	4.651	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
32) Benzene-d6	5.718	84	108332	4.480	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
36) 1,2-Dichloropropane-d6	6.682	67	32908	4.618	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.400%	
41) Toluene-d8	7.891	98	92944	4.339	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.800%	
43) trans-1,3-Dichloroprop...	8.174	79	9176	4.019	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	80.400%	
46) 2-Hexanone-d5	8.627	63	37104	40.124	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	80.240%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20838	4.488	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.800%	
66) 1,2-Dichlorobenzene-d4	12.187	152	29117	4.954	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
Data File : VU058265.D
Acq On : 29 Mar 2024 13:33
Operator : MD/SY
Sample : P1902-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOC4

Quant Time: Mar 29 23:42:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 29 23:40:26 2024
Response via : Initial Calibration

