

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
 Data File : VU058284.D
 Acq On : 29 Mar 2024 21:22
 Operator : MD/SY
 Sample : VU0329WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK158

Quant Time: Mar 30 01:09:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 30 01:09:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	69457	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	66941	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28686	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	29931	5.099	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.000%		
7) Chloroethane-d5	1.907	69	26520	5.535	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 110.800%		
11) 1,1-Dichloroethene-d2	2.560	65	12388	5.083	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 101.600%		
20) 2-Butanone-d5	4.627	46	55456	53.182	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 106.360%		
24) Chloroform-d	5.052	84	52898	5.041	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.800%		
26) 1,2-Dichloroethane-d4	5.695	65	25593	5.217	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.400%		
32) Benzene-d6	5.717	84	106863	4.915	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.200%		
36) 1,2-Dichloropropane-d6	6.682	67	33448	5.220	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.400%		
41) Toluene-d8	7.891	98	90529	4.700	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 94.000%		
43) trans-1,3-Dichloroprop...	8.177	79	9080	4.422	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 88.400%		
46) 2-Hexanone-d5	8.627	63	40073	48.193	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 96.380%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	21083	5.050	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 101.000%		
66) 1,2-Dichlorobenzene-d4	12.187	152	28450	5.326	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.600%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
Data File : VU058284.D
Acq On : 29 Mar 2024 21:22
Operator : MD/SY
Sample : VU0329WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK158

Quant Time: Mar 30 01:09:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 30 01:09:11 2024
Response via : Initial Calibration

