

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
 Data File : VU058287.D
 Acq On : 29 Mar 2024 22:33
 Operator : MD/SY
 Sample : P1912-07DL 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOD8DL

Quant Time: Mar 30 01:34:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 30 01:09:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	70350	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	67969	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	29473	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	18418	3.098	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.000%
7) Chloroethane-d5	1.908	69	20025	4.127	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	2.560	65	8013	3.246	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	4.628	46	61612	58.335	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.680%
24) Chloroform-d	5.052	84	47056	4.427	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
26) 1,2-Dichloroethane-d4	5.692	65	23732	4.776	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
32) Benzene-d6	5.718	84	87995	3.986	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.800%
36) 1,2-Dichloropropane-d6	6.682	67	29920	4.598	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.000%
41) Toluene-d8	7.891	98	69982	3.578	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.600%
43) trans-1,3-Dichloroprop...	8.178	79	8133	3.901	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.628	63	43614	51.658	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.320%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	21608	5.098	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	25929	4.724	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
Data File : VU058287.D
Acq On : 29 Mar 2024 22:33
Operator : MD/SY
Sample : P1912-07DL 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOD8DL

Quant Time: Mar 30 01:34:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 30 01:09:11 2024
Response via : Initial Calibration

