

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
 Data File : VU058288.D
 Acq On : 29 Mar 2024 22:56
 Operator : MD/SY
 Sample : P1912-08DL 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOD9DL

Quant Time: Mar 30 01:34:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 30 01:09:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	67552	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	64298	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	27655	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18049	3.162	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	63.200%	
7) Chloroethane-d5	1.907	69	19181	4.116	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.400%	
11) 1,1-Dichloroethene-d2	2.560	65	8042	3.393	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.800%	
20) 2-Butanone-d5	4.631	46	61339	60.482	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.960%	
24) Chloroform-d	5.052	84	45250	4.434	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.600%	
26) 1,2-Dichloroethane-d4	5.692	65	23811	4.990	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.800%	
32) Benzene-d6	5.717	84	85999	4.118	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.400%	
36) 1,2-Dichloropropane-d6	6.682	67	29379	4.773	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.400%	
41) Toluene-d8	7.891	98	67547	3.651	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.000%	
43) trans-1,3-Dichloroprop...	8.174	79	7806	3.958	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.200%	
46) 2-Hexanone-d5	8.631	63	42025	52.618	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	105.240%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	21423	5.343	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.800%	
66) 1,2-Dichlorobenzene-d4	12.187	152	23619	4.586	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032924\
Data File : VU058288.D
Acq On : 29 Mar 2024 22:56
Operator : MD/SY
Sample : P1912-08DL 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOD9DL

Quant Time: Mar 30 01:34:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 30 01:09:11 2024
Response via : Initial Calibration

