

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU033019\
 Data File : VU030547.D
 Acq On : 29 Mar 2019 17:21
 Operator : JC/SP
 Sample : VU0330WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK34

Quant Time: Mar 30 02:30:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 30 02:12:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	341999	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	335840	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	133204	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	144896	52.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.28%
7) Chloroethane-d5	1.67	69	118667	54.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.78%
11) 1,1-Dichloroethene-d2	2.27	63	210077	40.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.48%
21) 2-Butanone-d5	4.18	46	270505	104.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.55%
24) Chloroform-d	4.65	84	258885	54.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.18%
26) 1,2-Dichloroethane-d4	5.31	65	181583	58.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.04%
32) Benzene-d6	5.34	84	521472	54.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.30%
36) 1,2-Dichloropropane-d6	6.33	67	187224	56.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.64%
41) Toluene-d8	7.56	98	450592	52.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.18%
43) trans-1,3-Dichloropropene-	7.85	79	80261	51.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.60%
47) 2-Hexanone-d5	8.31	63	166547	93.27	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.27%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	253737	53.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.62%
64) 1,2-Dichlorobenzene-d4	11.86	152	152006	56.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU033019\
Data File : VU030547.D
Acq On : 29 Mar 2019 17:21
Operator : JC/SP
Sample : VU0330WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK34

Quant Time: Mar 30 02:30:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 30 02:12:04 2019
Response via : Initial Calibration

