

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033021\
 Data File : VU042839.D
 Acq On : 30 Mar 2021 15:18
 Operator : SY/MD
 Sample : M1846-22DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAR15DL

Quant Time: Mar 31 02:11:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 31 02:07:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	80024	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	80455	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	38333	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	19788	3.71	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.20%	
7) Chloroethane-d5	1.919	69	18443	4.57	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.40%	
11) 1,1-Dichloroethene-d2	2.575	63	35849	2.98	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	59.60%#	
20) 2-Butanone-d5	4.649	46	69834	39.87	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.74%	
24) Chloroform-d	5.073	84	48060	4.56	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.20%	
26) 1,2-Dichloroethane-d4	5.713	65	30534	4.70	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.00%	
32) Benzene-d6	5.739	84	85011	4.31	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.20%	
36) 1,2-Dichloropropane-d6	6.703	67	28732	4.52	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.40%	
41) Toluene-d8	7.909	98	76585	3.98	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.60%	
43) trans-1,3-Dichloroprop...	8.189	79	11377	3.83	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.60%	
46) 2-Hexanone-d5	8.648	63	64325	45.19	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.38%	
57) 1,1,2,2-Tetrachloroeth...	10.767	84	24737	4.36	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.20%	
64) 1,2-Dichlorobenzene-d4	12.201	152	32156	4.63	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.60%	
Target Compounds					Qvalue	
13) Acetone	2.665	43	3483	2.58	ug/L	96
34) Trichloroethene	6.552	95	43124	6.18	ug/L	99
47) Tetrachloroethene	8.562	164	6552	1.15	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033021\
Data File : VU042839.D
Acq On : 30 Mar 2021 15:18
Operator : SY/MD
Sample : M1846-22DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAR15DL

Quant Time: Mar 31 02:11:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 31 02:07:34 2021
Response via : Initial Calibration

