

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033119\
 Data File : VU030572.D
 Acq On : 30 Mar 2019 13:49
 Operator : JC/SP
 Sample : I.BLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK35

Quant Time: Mar 31 01:16:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U033119DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Sun Mar 31 01:11:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	49220	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	15433	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	16040	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	
Target Compounds					Ovalue	
3) Chloromethane	1.33	50	9634	0.360	ug/l	98
10) Iodomethane	2.41	142	696	0.543	ug/l #	80
13) Acetone	2.32	43	2895	0.838	ug/l #	79
18) 2-Butanone	4.31	43	946	0.212	ug/l	79
24) tert-Butyl Alcohol	2.84	59	328	0.257	ug/l #	88
45) 4-Methyl-2-Pentanone	7.47	43	6741	0.641	ug/l #	96
46) t-1,4-Dichloro-2-butene	10.31	75	7400	2.192	ug/l #	24
54) 2-Hexanone	8.38	43	10536	1.483	ug/l	87
71) 1,2,3-Trichloropropane	10.31	75	7400	0.795	ug/l #	41
80) Nitrobenzene	12.94	77	50	0.254	ug/l #	43
87) 1,2,4-Trichlorobenzene	13.51	180	4734	0.300	ug/l	98
89) Naphthalene	13.75	128	16347	0.594	ug/l #	96
90) 1,2,3-Trichlorobenzene	13.99	180	5681	0.379	ug/l	89
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033119\
Data File : VU030572.D
Acq On : 30 Mar 2019 13:49
Operator : JC/SP
Sample : I.BLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK35

Quant Time: Mar 31 01:16:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U033119DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Sun Mar 31 01:11:51 2019
Response via : Initial Calibration

