

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033119\
 Data File : VU030576.D
 Acq On : 30 Mar 2019 17:00
 Operator : JC/SP
 Sample : VU0331WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0331WBL01

Quant Time: Mar 31 01:38:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U033119DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Sun Mar 31 01:11:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	43992	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	15074	0.90	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	14299	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	

Target Compounds	Ovalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU033119\
Data File : VU030576.D
Acq On : 30 Mar 2019 17:00
Operator : JC/SP
Sample : VU0331WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0331WBL01

Quant Time: Mar 31 01:38:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U033119DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Sun Mar 31 01:11:51 2019
Response via : Initial Calibration

