


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU033119\
 Data File : VU030565.D
 Acq On : 30 Mar 2019 09:23
 Operator : JC/SP
 Sample : BFB
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U033119DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Sun Mar 31 01:11:51 2019



AutoFind: Scans 2865, 2866, 2867; Background Corrected with Scan 2851

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	95096	PASS
75	95	30	60	52.3	249749	PASS
95	95	100	100	100.0	477781	PASS
96	95	5	9	6.5	30952	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.1	349354	PASS
175	174	5	9	7.9	27680	PASS
176	174	95	101	97.5	340565	PASS
177	176	5	9	6.4	21933	PASS