

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033120\
 Data File : VU037503.D
 Acq On : 31 Mar 2020 18:16
 Operator : JC/MD
 Sample : L2063-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFCJ7

Quant Time: Apr 01 01:48:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR033120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 01 01:25:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	123056	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	116275	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	54098	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	22071	2.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	49.40%
7) Chloroethane-d5	1.93	69	21363	2.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	54.80%#
11) 1,1-Dichloroethene-d2	2.59	63	32113	1.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	36.60%#
20) 2-Butanone-d5	4.69	46	88013	31.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	63.58%
24) Chloroform-d	5.10	84	47792	2.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	55.80%#
26) 1,2-Dichloroethane-d4	5.74	65	27462	2.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	57.20%#
32) Benzene-d6	5.76	84	91007	2.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	53.40%#
36) 1,2-Dichloropropane-d6	6.72	67	32789	2.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	58.80%#
41) Toluene-d8	7.92	98	79630	2.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	50.60%#
43) trans-1,3-Dichloropropene-	8.20	79	13089	2.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	52.60%#
46) 2-Hexanone-d5	8.66	63	74875	32.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	64.94%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	27279	2.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	58.60%#
64) 1,2-Dichlorobenzene-d4	12.21	152	29343	2.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	56.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033120\
Data File : VU037503.D
Acq On : 31 Mar 2020 18:16
Operator : JC/MD
Sample : L2063-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFCJ7

Quant Time: Apr 01 01:48:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR033120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 01 01:25:39 2020
Response via : Initial Calibration

