

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033120\
 Data File : VU037511.D
 Acq On : 31 Mar 2020 21:36
 Operator : JC/MD
 Sample : L2065-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBF04

Quant Time: Apr 01 03:25:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR033120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 01 01:25:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	89285	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	82229	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	38366	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	22343	3.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.00%
7) Chloroethane-d5	1.93	69	21895	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.40%
11) 1,1-Dichloroethene-d2	2.59	63	31426	2.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.40%#
20) 2-Butanone-d5	4.69	46	86642	43.13	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.26%
24) Chloroform-d	5.11	84	48567	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.00%
26) 1,2-Dichloroethane-d4	5.74	65	27507	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
32) Benzene-d6	5.76	84	91879	3.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.40%
36) 1,2-Dichloropropane-d6	6.72	67	33663	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.92	98	80388	3.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.40%
43) trans-1,3-Dichloropropene-	8.21	79	13929	3.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.20%
46) 2-Hexanone-d5	8.66	63	80514	49.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.74%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	29600	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.00%
64) 1,2-Dichlorobenzene-d4	12.20	152	30819	4.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.00%

Target Compounds					Ovalue
25) Chloroform	5.12	83	3316	0.277 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU033120\
Data File : VU037511.D
Acq On : 31 Mar 2020 21:36
Operator : JC/MD
Sample : L2065-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBF04

Quant Time: Apr 01 03:25:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR033120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 01 01:25:39 2020
Response via : Initial Calibration

