

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033121\
 Data File : VU042861.D
 Acq On : 31 Mar 2021 16:40
 Operator : SY/MD
 Sample : M1828-04
 Misc : 6.94g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FS-SB14(5-6)

Quant Time: Apr 01 04:43:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	125769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	200835	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.426	117	197042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	107965	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	99442	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
35) Dibromofluoromethane	5.298	113	68234	45.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.04%	
50) Toluene-d8	7.906	98	251822	47.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.52%	
62) 4-Bromofluorobenzene	10.645	95	99684	48.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.30%	
Target Compounds						
95) Naphthalene	14.095	128	85050	13.60	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033121\
Data File : VU042861.D
Acq On : 31 Mar 2021 16:40
Operator : SY/MD
Sample : M1828-04
Misc : 6.94g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FS-SB14(5-6)

Quant Time: Apr 01 04:43:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

