

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033121\
 Data File : VU042862.D
 Acq On : 31 Mar 2021 17:03
 Operator : SY/MD
 Sample : M1836-05ME
 Misc : 6.49g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B3-0-2ME

Quant Time: Apr 01 04:43:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	123923	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	198116	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.426	117	200014	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	111724	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	97305	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	5.295	113	66532	45.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.00%	
50) Toluene-d8	7.906	98	249229	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
62) 4-Bromofluorobenzene	10.645	95	104279	51.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.18%	
Target Compounds						
						Qvalue
64) Tetrachloroethene	8.558	164	1953	1.21	ug/l #	87
95) Naphthalene	14.095	128	43620	8.04	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033121\
Data File : VU042862.D
Acq On : 31 Mar 2021 17:03
Operator : SY/MD
Sample : M1836-05ME
Misc : 6.49g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B3-0-2ME

Quant Time: Apr 01 04:43:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

