

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033122\
 Data File : VU047773.D
 Acq On : 31 Mar 2022 15:12
 Operator : SY/MD
 Sample : VU0331WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0331WBL01

Quant Time: Apr 01 02:24:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U033122W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 13:50:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	322428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	488151	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	475150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	257129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	197362	57.057	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.120%
35) Dibromofluoromethane	5.292	113	178016	54.378	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.760%
50) Toluene-d8	7.899	98	617427	52.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.700%
62) 4-Bromofluorobenzene	10.636	95	227356	49.498	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033122\
Data File : VU047773.D
Acq On : 31 Mar 2022 15:12
Operator : SY/MD
Sample : VU0331WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0331WBL01

Quant Time: Apr 01 02:24:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U033122W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 13:50:29 2022
Response via : Initial Calibration

