

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033122\
 Data File : VU047774.D
 Acq On : 31 Mar 2022 15:35
 Operator : SY/MD
 Sample : VU0331WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0331WBL02

Quant Time: Apr 01 02:25:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U033122W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 13:50:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	316475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	481653	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	463392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	247975	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	190287	56.046	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.100%
35) Dibromofluoromethane	5.292	113	170744	52.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.720%
50) Toluene-d8	7.899	98	586299	50.382	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.760%
62) 4-Bromofluorobenzene	10.632	95	214106	47.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU033122\
Data File : VU047774.D
Acq On : 31 Mar 2022 15:35
Operator : SY/MD
Sample : VU0331WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0331WBL02

Quant Time: Apr 01 02:25:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U033122W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 13:50:29 2022
Response via : Initial Calibration

