

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040120\
 Data File : VU037539.D
 Acq On : 01 Apr 2020 22:54
 Operator : JC/MD
 Sample : L2065-18
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBF14

Quant Time: Apr 02 03:34:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 02 01:37:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.34	114	78892	5.00	ug/L	0.05
28) Chlorobenzene-d5	9.45	117	74121	5.00	ug/L	0.01
60) 1,4-Dichlorobenzene-d4	11.83	152	39810	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.66	65	20547	3.11	ug/L	0.05
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.20%
7) Chloroethane-d5	1.99	69	18247	3.59	ug/L	0.06
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.80%
11) 1,1-Dichloroethene-d2	2.64	63	29278	2.49	ug/L	0.05
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.80%#
20) 2-Butanone-d5	4.81	46	74418	41.50	ug/L	0.13
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.00%
24) Chloroform-d	5.18	84	44707	3.95	ug/L	0.08
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
26) 1,2-Dichloroethane-d4	5.81	65	24033	3.84	ug/L	0.07
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
32) Benzene-d6	5.83	84	84847	3.71	ug/L	0.07
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.20%
36) 1,2-Dichloropropane-d6	6.77	67	29618	3.95	ug/L	0.05
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.00%
41) Toluene-d8	7.95	98	71865	3.43	ug/L	0.03
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.60%#
43) trans-1,3-Dichloropropene-	8.23	79	11360	3.36	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.20%
46) 2-Hexanone-d5	8.68	63	70163	42.91	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.82%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	25546	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	31733	3.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.40%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.81	43	16340	16.164	ug/L	42
14) Carbon disulfide	2.87	76	12181	0.711	ug/L #	89
25) Chloroform	5.20	83	42801	3.906	ug/L	98
33) Benzene	5.88	78	72376	3.076	ug/L	100
38) Bromodichloromethane	7.17	83	3743	0.462	ug/L	94
42) Toluene	8.02	91	502203	20.082	ug/L	96
52) Ethylbenzene	9.60	91	641188	22.503	ug/L	100
53) m,p-Xylene	9.72	106	580829	54.232	ug/L	95
54) o-Xylene	10.13	106	231697	22.462	ug/L	99
55) Styrene	10.14	104	344865	19.357	ug/L	98
56) Isopropylbenzene	10.51	105	149619	5.326	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040120\
Data File : VU037539.D
Acq On : 01 Apr 2020 22:54
Operator : JC/MD
Sample : L2065-18
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBF14

Quant Time: Apr 02 03:34:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 02 01:37:22 2020
Response via : Initial Calibration

