

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040120\
 Data File : VU037543.D
 Acq On : 02 Apr 2020 00:33
 Operator : JC/MD
 Sample : L2064-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 02 04:31:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 02 01:37:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	85407	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	81198	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	40884	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	29954	4.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.80%
7) Chloroethane-d5	1.93	69	25624	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	2.59	63	44111	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	4.67	46	87997	45.33	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.66%
24) Chloroform-d	5.11	84	52960	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	5.74	65	31178	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	5.76	84	111390	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.72	67	37307	4.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.80%
41) Toluene-d8	7.92	98	104287	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	8.20	79	15764	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
46) 2-Hexanone-d5	8.66	63	83258	46.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.96%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	30416	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	12.21	152	37471	4.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.00%

Target Compounds					Ovalue
25) Chloroform	5.12	83	2546	0.215 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040120\
Data File : VU037543.D
Acq On : 02 Apr 2020 00:33
Operator : JC/MD
Sample : L2064-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 02 04:31:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 02 01:37:22 2020
Response via : Initial Calibration

