

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
 Data File : VU042873.D
 Acq On : 01 Apr 2021 12:43
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Apr 01 14:35:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 01 14:29:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	110778	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	108264	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	54398	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31608	4.58	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	91.60%		
7) Chloroethane-d5	1.919	69	27427	4.56	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	91.20%		
11) 1,1-Dichloroethene-d2	2.575	63	57680	3.53	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	70.60%		
20) 2-Butanone-d5	4.649	46	118265	42.40	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	84.80%		
24) Chloroform-d	5.076	84	71421	4.61	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	92.20%		
26) 1,2-Dichloroethane-d4	5.713	65	47185	4.60	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	92.00%		
32) Benzene-d6	5.739	84	133837	4.78	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	95.60%		
36) 1,2-Dichloropropane-d6	6.700	67	42522	4.63	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	92.60%		
41) Toluene-d8	7.909	98	123185	4.43	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	88.60%		
43) trans-1,3-Dichloroprop...	8.189	79	18501	4.52	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	90.40%		
46) 2-Hexanone-d5	8.648	63	100201	48.03	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	96.06%		
57) 1,1,2,2-Tetrachloroeth...	10.767	84	37550	4.59	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	91.80%		
64) 1,2-Dichlorobenzene-d4	12.205	152	49935	4.90	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	98.00%		
Target Compounds						
3) Chloromethane	1.523	50	408	0.05	ug/L	86
10) 1,1,2-Trichloro-1,2,2-...	2.575	101	521	0.07	ug/L #	18
12) 1,1-Dichloroethene	2.575	96	461	0.06	ug/L #	1
14) Carbon disulfide	2.797	76	1208	0.05	ug/L #	75
16) Methylene chloride	3.057	84	1373	0.17	ug/L	95
35) Methylcyclohexane	6.700	83	10028	0.79	ug/L #	15
37) 1,2-Dichloropropane	6.703	63	5450	0.64	ug/L #	87
39) cis-1,3-Dichloropropene	7.574	75	1609	0.13	ug/L	85
44) trans-1,3-Dichloropropene	8.192	75	963	0.08	ug/L	83
48) 2-Hexanone	8.700	43	5088	0.92	ug/L #	68
63) 1,4-Dichlorobenzene	11.848	146	906	0.05	ug/L	86
65) 1,2-Dichlorobenzene	12.217	146	974	0.06	ug/L #	61
66) 1,2-Dibromo-3-chloropr...	13.002	75	83	0.06	ug/L #	1
67) 1,3,5-Trichlorobenzene	13.227	180	1056	0.07	ug/L	88
68) 1,2,4-trichlorobenzene	13.851	180	1732	0.15	ug/L	90
69) Naphthalene	14.095	128	5205	0.28	ug/L	98
70) 1,2,3-Trichlorobenzene	14.339	180	1982	0.19	ug/L	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
Data File : VU042873.D
Acq On : 01 Apr 2021 12:43
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Apr 01 14:35:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 01 14:29:07 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
Data File : VU042873.D
Acq On : 01 Apr 2021 12:43
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Apr 01 14:35:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 01 14:29:07 2021
Response via : Initial Calibration

