

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
 Data File : VU042884.D
 Acq On : 01 Apr 2021 18:50
 Operator : SY/MD
 Sample : M1847-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAR16

Quant Time: Apr 02 05:20:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 02 05:16:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	84012	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	83051	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	39591	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	24889	4.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.20%
7) Chloroethane-d5	1.919	69	22881	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.575	63	41035	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	4.645	46	108761	51.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.82%
24) Chloroform-d	5.073	84	55453	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.713	65	36118	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	5.735	84	102462	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
36) 1,2-Dichloropropane-d6	6.700	67	33971	4.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.60%
41) Toluene-d8	7.906	98	94188	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
43) trans-1,3-Dichloroprop...	8.185	79	13535	4.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.20%
46) 2-Hexanone-d5	8.645	63	78178	48.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.70%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	28955	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	12.201	152	38324	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%
Target Compounds					Qvalue	
30) Cyclohexane	5.401	56	3850	0.38	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
Data File : VU042884.D
Acq On : 01 Apr 2021 18:50
Operator : SY/MD
Sample : M1847-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAR16

Quant Time: Apr 02 05:20:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 02 05:16:37 2021
Response via : Initial Calibration

