

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
 Data File : VU042889.D
 Acq On : 02 Apr 2021 09:54
 Operator : SY/MD
 Sample : VU0401WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK06

Quant Time: Apr 02 10:13:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 02 05:16:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	79732	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	80568	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	38141	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	24297	4.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.80%
7) Chloroethane-d5	1.919	69	22349	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	2.575	63	41912	3.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.40%
20) 2-Butanone-d5	4.649	46	88232	43.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.90%
24) Chloroform-d	5.073	84	56169	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.713	65	36550	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.735	84	100116	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.700	67	34468	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.906	98	92291	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
43) trans-1,3-Dichloroprop...	8.185	79	13807	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.645	63	79461	51.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.36%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	29478	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
64) 1,2-Dichlorobenzene-d4	12.201	152	38112	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040121\
Data File : VU042889.D
Acq On : 02 Apr 2021 09:54
Operator : SY/MD
Sample : VU0401WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK06

Quant Time: Apr 02 10:13:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR040121WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 02 05:16:37 2021
Response via : Initial Calibration

