

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040122\
 Data File : VU047784.D
 Acq On : 01 Apr 2022 13:03
 Operator : SY/MD
 Sample : VU0401WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK077

Quant Time: Apr 02 01:07:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 02 01:06:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	348743	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	362259	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	199019	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	130498	51.601	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.200%
7) Chloroethane-d5	1.916	69	108363	54.342	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.680%
11) 1,1-Dichloroethene-d2	2.571	63	172835	37.354	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.700%
21) 2-Butanone-d5	4.636	46	182595	84.119	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.120%
24) Chloroform-d	5.067	84	238656	51.575	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.160%
26) 1,2-Dichloroethane-d4	5.703	65	155557	54.812	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.620%
32) Benzene-d6	5.729	84	492260	54.003	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.000%
36) 1,2-Dichloropropane-d6	6.693	67	154727	55.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.540%
41) Toluene-d8	7.899	98	454819	51.212	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.420%
43) trans-1,3-Dichloroprop...	8.179	79	76013	53.596	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	107.200%
47) 2-Hexanone-d5	8.636	63	140513	85.742	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.740%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	228801	47.624	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.240%
66) 1,2-Dichlorobenzene-d4	12.195	152	193807	54.736	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040122\
Data File : VU047784.D
Acq On : 01 Apr 2022 13:03
Operator : SY/MD
Sample : VU0401WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK077

Quant Time: Apr 02 01:07:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 02 01:06:35 2022
Response via : Initial Calibration

