

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040122\
 Data File : VU047785.D
 Acq On : 01 Apr 2022 13:32
 Operator : SY/MD
 Sample : N2179-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0R37

Quant Time: Apr 02 01:07:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 02 01:06:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	365627	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	368996	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	200342	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	140585	53.023	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.040%		
7) Chloroethane-d5	1.919	69	113576	54.326	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 108.660%		
11) 1,1-Dichloroethene-d2	2.571	63	186930	38.535	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 77.060%		
21) 2-Butanone-d5	4.639	46	173281	76.142	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 76.140%		
24) Chloroform-d	5.067	84	248456	51.214	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.420%		
26) 1,2-Dichloroethane-d4	5.706	65	161343	54.225	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.460%		
32) Benzene-d6	5.732	84	515756	55.547	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 111.100%		
36) 1,2-Dichloropropane-d6	6.693	67	160607	56.324	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 112.640%		
41) Toluene-d8	7.899	98	483256	53.420	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.840%		
43) trans-1,3-Dichloroprop...	8.182	79	77068	53.348	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 106.700%		
47) 2-Hexanone-d5	8.639	63	132111	79.143	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 79.140%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	225490	46.078	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 92.160%		
66) 1,2-Dichlorobenzene-d4	12.195	152	198469	55.683	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 111.360%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040122\
Data File : VU047785.D
Acq On : 01 Apr 2022 13:32
Operator : SY/MD
Sample : N2179-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0R37

Quant Time: Apr 02 01:07:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 02 01:06:35 2022
Response via : Initial Calibration

