

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040124\
 Data File : VU058320.D
 Acq On : 01 Apr 2024 17:51
 Operator : MD/SY
 Sample : P1912-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCOD9

Quant Time: Apr 02 02:29:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 02 02:24:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	69504	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	67369	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	30466	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	24832	4.228	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.600%
7) Chloroethane-d5	1.908	69	23832	4.971	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.560	65	10794	4.426	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.600%
20) 2-Butanone-d5	4.621	46	51994	49.828	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.660%
24) Chloroform-d	5.055	84	48779	4.645	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.695	65	24390	4.968	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.718	84	100211	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.682	67	31764	4.925	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.891	98	84718	4.370	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.174	79	8641	4.182	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.628	63	39141	46.773	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.540%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	20436	4.864	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	27603	4.865	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040124\
Data File : VU058320.D
Acq On : 01 Apr 2024 17:51
Operator : MD/SY
Sample : P1912-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCOD9

Quant Time: Apr 02 02:29:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 02 02:24:47 2024
Response via : Initial Calibration

