

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040220\
 Data File : VU037556.D
 Acq On : 02 Apr 2020 15:51
 Operator : JC/MD
 Sample : L2065-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 03 01:52:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 03 00:57:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	91732	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	85789	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	42876	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	32296	4.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.00%
7) Chloroethane-d5	1.93	69	27033	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.59	63	44923	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	4.66	46	92238	44.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.46%
24) Chloroform-d	5.11	84	55939	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	5.74	65	32432	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	5.76	84	117314	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.72	67	38273	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.20%
41) Toluene-d8	7.92	98	107142	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	8.20	79	17625	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.65	63	87193	46.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.14%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	31433	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	12.21	152	39556	4.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040220\
Data File : VU037556.D
Acq On : 02 Apr 2020 15:51
Operator : JC/MD
Sample : L2065-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 03 01:52:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR040120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 03 00:57:17 2020
Response via : Initial Calibration

