

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
 Data File : VU042901.D
 Acq On : 02 Apr 2021 17:23
 Operator : SY/MD
 Sample : M1903-02 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 30791

Quant Time: Apr 03 00:13:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	109663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	175892	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	170843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	85445	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	87071	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
35) Dibromofluoromethane	5.298	113	61857	47.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.24%	
50) Toluene-d8	7.906	98	221525	47.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.94%	
62) 4-Bromofluorobenzene	10.642	95	81304	45.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
Data File : VU042901.D
Acq On : 02 Apr 2021 17:23
Operator : SY/MD
Sample : M1903-02 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
30791

Quant Time: Apr 03 00:13:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

