

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
 Data File : VU042917.D
 Acq On : 02 Apr 2021 23:39
 Operator : SY/MD
 Sample : M1927-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-6

Quant Time: Apr 03 00:19:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	126107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	202650	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	192517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	95651	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	100468	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
35) Dibromofluoromethane	5.298	113	69701	46.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.18%	
50) Toluene-d8	7.906	98	254155	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	
62) 4-Bromofluorobenzene	10.639	95	94429	45.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.34%	
Target Compounds						
78) n-propylbenzene	10.912	91	6219	0.87	ug/l	97
95) Naphthalene	14.098	128	2007	2.87	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
Data File : VU042917.D
Acq On : 02 Apr 2021 23:39
Operator : SY/MD
Sample : M1927-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-6

Quant Time: Apr 03 00:19:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

