

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
 Data File : VU042919.D
 Acq On : 03 Apr 2021 00:25
 Operator : SY/MD
 Sample : M1903-07 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 31029

Quant Time: Apr 03 01:42:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	115885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	185534	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	177543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	86867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	92494	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	
35) Dibromofluoromethane	5.298	113	64310	46.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.88%	
50) Toluene-d8	7.906	98	236218	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
62) 4-Bromofluorobenzene	10.639	95	86038	45.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040221\
Data File : VU042919.D
Acq On : 03 Apr 2021 00:25
Operator : SY/MD
Sample : M1903-07 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
31029

Quant Time: Apr 03 01:42:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

