

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040320\
 Data File : VU037568.D
 Acq On : 03 Apr 2020 15:07
 Operator : JC/MD
 Sample : VU0403WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK46

Quant Time: Apr 04 05:38:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 04 05:18:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	176448	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	165907	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	79168	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	65136	45.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.90%
7) Chloroethane-d5	1.93	69	51656	44.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.72%
11) 1,1-Dichloroethene-d2	2.59	63	89667	37.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.44%
21) 2-Butanone-d5	4.66	46	108712	100.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.93%
24) Chloroform-d	5.11	84	113141	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.72%
26) 1,2-Dichloroethane-d4	5.74	65	74349	46.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.76%
32) Benzene-d6	5.76	84	229045	47.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.30%
36) 1,2-Dichloropropane-d6	6.72	67	78050	48.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.46%
41) Toluene-d8	7.92	98	209141	46.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.36%
43) trans-1,3-Dichloropropene-	8.20	79	40012	45.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.08%
47) 2-Hexanone-d5	8.65	63	84187	98.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.20%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	109919	47.52	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.04%
64) 1,2-Dichlorobenzene-d4	12.21	152	77339	47.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040320\
Data File : VU037568.D
Acq On : 03 Apr 2020 15:07
Operator : JC/MD
Sample : VU0403WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK46

Quant Time: Apr 04 05:38:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 04 05:18:00 2020
Response via : Initial Calibration

