

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040323\
 Data File : VU053512.D
 Acq On : 03 Apr 2023 12:06
 Operator : JC/MD
 Sample : VU0403WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK064

Quant Time: Apr 04 05:16:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 01 04:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	262878	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	254612	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	131881	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	102632	50.139	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.280%		
7) Chloroethane-d5	1.909	69	83395	52.841	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 105.680%		
11) 1,1-Dichloroethene-d2	2.565	63	121167	38.240	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 76.480%		
21) 2-Butanone-d5	4.613	46	105945	114.318	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 114.320%		
24) Chloroform-d	5.057	84	160607	47.443	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 94.880%		
26) 1,2-Dichloroethane-d4	5.697	65	103224	51.603	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.200%		
32) Benzene-d6	5.723	84	333958	48.592	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 97.180%		
36) 1,2-Dichloropropane-d6	6.684	67	104549	49.769	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 99.540%		
41) Toluene-d8	7.893	98	313974	45.964	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 91.920%		
43) trans-1,3-Dichloroprop...	8.173	79	51794	47.765	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 95.540%		
47) 2-Hexanone-d5	8.626	63	83370	116.288	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 116.290%		
56) 1,1,2,2-Tetrachloroeth...	10.751	84	155374	52.478	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 104.960%		
66) 1,2-Dichlorobenzene-d4	12.188	152	129885	47.182	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 94.360%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040323\
Data File : VU053512.D
Acq On : 03 Apr 2023 12:06
Operator : JC/MD
Sample : VU0403WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK064

Quant Time: Apr 04 05:16:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 01 04:51:17 2023
Response via : Initial Calibration

