




Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU040419\
 Data File : VU030648.D
 Acq On : 03 Apr 2019 23:17
 Operator : JC/SP
 Sample : K2167-12ME
 Misc : 4.70µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Quant Time: Apr 04 06:59:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 06:52:42 2019
 Response via : Initial Calibration



TIC: VU030648.D

(7) Chloroethane-d5 (S)

1.685min (+0.010) 0.16ug/L

response 535

Ion	Exp%	Act%
69.00	100	100
71.00	31.70	0.00#
51.00	28.60	0.00#
0.00	0.00	0.00