

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU040419\
 Data File : VU030626.D
 Acq On : 03 Apr 2019 14:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05044

Quant Time: Apr 04 06:51:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 02 08:36:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.552	0.441	20.1	83	0.00
3 T	Chloromethane	0.617	0.557	9.7	94	0.00
4 S	Vinyl Chloride-d3	0.402	0.429	-6.7	112	0.00
5 T	Vinyl chloride	0.585	0.528	9.7	94	0.00
6 T	Bromomethane	0.268	0.279	-4.1	109	0.00
7 S	Chloroethane-d5	0.319	0.334	-4.7	109	0.00
8 T	Chloroethane	0.330	0.306	7.3	97	0.00
9 T	Trichlorofluoromethane	0.638	0.625	2.0	101	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.365	0.354	3.0	100	0.00
11 S	1,1-Dichloroethene-d2	0.763	0.797	-4.5	108	0.00
12 T	1,1-Dichloroethene	0.358	0.337	5.9	98	0.00
13 T	Acetone	0.383	0.329	14.1	86	0.00
14 T	Carbon disulfide	1.183	1.086	8.2	97	0.00
15 T	Methyl Acetate	0.586	0.515	12.1	88	0.00
16 T	Methylene chloride	0.422	0.390	7.6	97	0.00
17 T	trans-1,2-Dichloroethene	0.375	0.346	7.7	95	0.00
18 T	Methyl tert-butyl Ether	1.241	1.099	11.4	91	0.00
19 T	1,1-Dichloroethane	0.770	0.715	7.1	97	0.00
20 T	cis-1,2-Dichloroethene	0.421	0.381	9.5	93	0.00
21 S	2-Butanone-d5	0.378	0.307	18.8	79	0.00
22 T	2-Butanone	0.462	0.405	12.3	84	0.00
23 T	Bromochloromethane	0.196	0.185	5.6	98	0.00
24 S	Chloroform-d	0.693	0.655	5.5	95	0.00
25 T	Chloroform	0.732	0.706	3.6	100	0.00
26 S	1,2-Dichloroethane-d4	0.454	0.433	4.6	98	0.00
27 T	1,2-Dichloroethane	0.574	0.546	4.9	97	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
29 T	Cyclohexane	0.757	0.690	8.9	92	0.00
30 T	1,1,1-Trichloroethane	0.612	0.580	5.2	99	0.00
31 T	Carbon tetrachloride	0.519	0.513	1.2	101	0.00
32 S	Benzene-d6	1.434	1.361	5.1	95	0.00
33 T	Benzene	1.702	1.589	6.6	96	0.00
34 T	Trichloroethene	0.418	0.381	8.9	95	0.00
35 T	Methylcyclohexane	0.689	0.614	10.9	93	0.00
36 S	1,2-Dichloropropane-d6	0.491	0.460	6.3	94	0.00
37 T	1,2-Dichloropropane	0.483	0.440	8.9	94	0.00
38 T	Bromodichloromethane	0.574	0.543	5.4	98	0.00
39 T	cis-1,3-Dichloropropene	0.694	0.648	6.6	93	0.00
40 T	4-Methyl-2-pentanone	0.811	0.687	15.3	86	0.00
41 S	Toluene-d8	1.288	1.219	5.4	94	0.00
42 T	Toluene	1.766	1.617	8.4	93	0.00
43 S	trans-1,3-Dichloropropene-d	0.231	0.213	7.8	94	0.00
44 T	trans-1,3-Dichloropropene	0.611	0.571	6.5	93	0.00
45 T	1,1,2-Trichloroethane	0.408	0.368	9.8	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU040419\
 Data File : VU030626.D
 Acq On : 03 Apr 2019 14:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05044

Quant Time: Apr 04 06:51:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 02 08:36:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Tetrachloroethene	0.292	0.275	5.8	96	0.00
47 S	2-Hexanone-d5	0.266	0.204	23.3	75	0.00
48 T	2-Hexanone	0.665	0.551	17.1	84	0.00
49 T	Dibromochloromethane	0.424	0.400	5.7	96	0.00
50 T	1,2-Dibromoethane	0.438	0.389	11.2	91	0.00
51 T	Chlorobenzene	1.071	0.968	9.6	94	0.00
52 T	Ethylbenzene	1.914	1.770	7.5	94	0.00
53 T	m,p-Xylene	0.691	0.636	8.0	95	0.00
54 T	o-xylene	0.674	0.617	8.5	92	0.00
55 T	Styrene	1.161	1.077	7.2	92	0.00
56 T	Isopropylbenzene	1.785	1.637	8.3	93	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.702	0.618	12.0	89	0.00
58 T	1,1,2,2-Tetrachloroethane	0.752	0.662	12.0	91	0.00
59	1,2,3-Trichloropropane	0.588	0.519	11.7	91	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
61 T	Bromoform	0.665	0.614	7.7	91	0.00
62 T	1,3-Dichlorobenzene	1.635	1.474	9.8	91	0.00
63 T	1,4-Dichlorobenzene	1.683	1.506	10.5	91	0.00
64 S	1,2-Dichlorobenzene-d4	1.002	0.930	7.2	93	0.00
65 T	1,2-Dichlorobenzene	1.653	1.519	8.1	92	0.00
66 T	1,2-Dibromo-3-chloropropane	0.368	0.317	13.9	86	0.00
67	1,3,5-Trichlorobenzene	1.224	1.139	6.9	92	0.00
68 T	1,2,4-trichlorobenzene	1.064	0.999	6.1	92	0.00
69	Naphthalene	3.782	3.183	15.8	87	0.00
70 T	1,2,3-Trichlorobenzene	1.123	1.041	7.3	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0