

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040419\
 Data File : VU030674.D
 Acq On : 04 Apr 2019 10:11
 Operator : JC/SP
 Sample : VU0404WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK40

Quant Time: Apr 04 13:27:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 13:18:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	516992	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	505382	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	223046	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	209325	50.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.62%
7) Chloroethane-d5	1.68	69	165984	50.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.66%
11) 1,1-Dichloroethene-d2	2.27	63	289662	36.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.40%
21) 2-Butanone-d5	4.17	46	377764	96.58	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.58%
24) Chloroform-d	4.64	84	322998	45.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.12%
26) 1,2-Dichloroethane-d4	5.31	65	219805	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.72%
32) Benzene-d6	5.33	84	684241	47.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.42%
36) 1,2-Dichloropropane-d6	6.33	67	239243	48.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.50%
41) Toluene-d8	7.56	98	603823	46.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.78%
43) trans-1,3-Dichloropropene-	7.85	79	103080	44.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.42%
47) 2-Hexanone-d5	8.31	63	240536	89.52	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.52%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	322065	45.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.78%
64) 1,2-Dichlorobenzene-d4	11.86	152	212952	47.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040419\
Data File : VU030674.D
Acq On : 04 Apr 2019 10:11
Operator : JC/SP
Sample : VU0404WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK40

Quant Time: Apr 04 13:27:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 13:18:40 2019
Response via : Initial Calibration

