

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040423\
 Data File : VU053556.D
 Acq On : 04 Apr 2023 12:56
 Operator : JC/MD
 Sample : 02158-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PMW-1(20230328)

Quant Time: Apr 04 13:16:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 01 04:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	223424	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	211140	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	105742	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	79103	45.469	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	90.940%	
7) Chloroethane-d5	1.909	69	71436	53.256	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	106.520%	
11) 1,1-Dichloroethene-d2	2.565	63	100301	37.245	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.480%	
21) 2-Butanone-d5	4.613	46	104427	132.577	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	132.580%#	
24) Chloroform-d	5.060	84	151276	52.578	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.160%	
26) 1,2-Dichloroethane-d4	5.697	65	97132	57.132	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	114.260%	
32) Benzene-d6	5.723	84	309047	54.226	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.460%	
36) 1,2-Dichloropropane-d6	6.687	67	96057	55.142	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	110.280%	
41) Toluene-d8	7.893	98	286537	50.584	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	101.160%	
43) trans-1,3-Dichloroprop...	8.176	79	48182	53.583	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	107.160%	
47) 2-Hexanone-d5	8.626	63	81321	136.784	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	136.780%#	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	140128	57.073	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	114.140%	
66) 1,2-Dichlorobenzene-d4	12.189	152	115894	52.507	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.020%	
Target Compounds						
					Qvalue	
51) Chlorobenzene	9.446	112	5580	1.299	ug/L #	86
64) 1,3-Dichlorobenzene	11.745	146	3195	0.978	ug/L	89
65) 1,4-Dichlorobenzene	11.832	146	9784	2.939	ug/L	95
70) 1,2,4-trichlorobenzene	13.841	180	5908	2.634	ug/L #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040423\
Data File : VU053556.D
Acq On : 04 Apr 2023 12:56
Operator : JC/MD
Sample : 02158-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PMW-1(20230328)

Quant Time: Apr 04 13:16:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 01 04:51:17 2023
Response via : Initial Calibration

