


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040423\
Data File : VU053550.D
Acq On : 04 Apr 2023 09:56
Operator : JC/MD
Sample : BFB173
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB173

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Title : VOC Analysis
Last Update : Sat Apr 01 04:51:17 2023



AutoFind: Scans 2888, 2889, 2890; Background Corrected with Scan 2877

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.9	6147	PASS
75	95	30	80	52.4	20277	PASS
95	95	100	100	100.0	38704	PASS
96	95	5	9	6.8	2622	PASS
173	174	0.00	2	0.8	260	PASS
174	95	50	120	81.4	31504	PASS
175	174	5	9	8.0	2520	PASS
176	174	95	101	98.1	30891	PASS
177	176	5	9	5.9	1823	PASS

m/z	Abundance
43.90	227.0

Instrument :
MSVOA_U
ClientSampleId :
BFB173

m/z	Abundance
36.20	180.0
37.10	1321.0
38.10	1147.0
39.10	642.0
44.10	532.0
45.10	309.0
47.00	363.0
47.90	219.0
49.10	1464.0
50.10	6151.0
51.10	2124.0
52.00	102.0
56.00	452.0
57.00	879.0
60.00	382.0
61.10	1677.0
62.10	1715.0
63.10	1128.0
64.10	199.0
65.00	192.0
68.10	3769.0
69.10	4100.0
70.10	394.0
72.00	149.0
73.10	1861.0
74.10	6343.0
75.10	20096.0
76.10	1949.0
77.10	293.0
77.90	160.0
78.20	152.0
79.00	1686.0
80.00	467.0
81.00	1799.0
82.00	355.0
87.00	1940.0
88.00	1758.0
90.80	155.0
92.10	1149.0
93.10	1392.0
94.10	4579.0
95.10	38768.0
96.10	2558.0
104.10	219.0
105.90	174.0
116.00	158.0
117.00	224.0
119.00	195.0
127.90	134.0
129.00	100.0
130.00	184.0
140.90	391.0
143.00	391.0
171.90	183.0
173.10	247.0
174.10	30160.0
175.00	2348.0
176.00	29808.0
177.00	1643.0

Instrument :
MSVOA_U
ClientSampleId :
BFB173

m/z	Abundance
35.90	153.0
36.30	174.0
37.10	1382.0
38.10	1131.0
39.10	612.0
44.10	524.0
45.10	327.0
47.00	368.0
47.90	203.0
49.10	1506.0
50.10	6277.0
51.10	2151.0
55.10	125.0
56.00	553.0
57.00	887.0
60.10	450.0
61.10	1746.0
62.10	1435.0
63.10	1304.0
64.10	172.0
65.00	200.0
67.00	100.0
68.10	3837.0
69.10	4415.0
70.10	439.0
72.00	161.0
73.10	1993.0
74.10	6568.0
75.10	21000.0
76.20	2101.0
77.00	225.0
78.10	217.0
79.00	1738.0
80.00	413.0
81.00	1841.0
81.90	372.0
87.00	2040.0
88.00	1934.0
90.80	150.0
92.10	1219.0
93.10	1571.0
94.10	4967.0
95.10	39920.0
96.10	2735.0
104.10	207.0
105.90	183.0
115.90	159.0
117.00	182.0
118.90	217.0
127.90	189.0
129.00	127.0
130.00	176.0
140.90	419.0
143.00	433.0
171.90	168.0
173.00	281.0
174.10	32368.0
175.00	2676.0
176.00	31800.0
177.00	1859.0

Instrument :
MSVOA_U
ClientSampleId :
BFB173

m/z	Abundance
35.80	189.0
36.30	182.0
37.10	1341.0
38.10	1144.0
39.20	642.0
44.10	461.0
45.10	281.0
47.10	339.0
48.00	209.0
49.10	1456.0
50.10	6014.0
51.10	2043.0
55.00	137.0
56.00	606.0
57.00	894.0
60.10	455.0
61.10	1677.0
62.10	1242.0
63.10	1298.0
64.20	135.0
65.00	159.0
68.10	3693.0
69.10	4049.0
70.10	419.0
72.00	149.0
73.10	1811.0
74.10	6110.0
75.10	19736.0
76.20	1979.0
76.90	217.0
78.10	242.0
78.90	1648.0
80.00	308.0
81.10	1699.0
81.90	376.0
87.00	1737.0
88.00	1810.0
90.80	116.0
92.10	1177.0
93.00	1641.0
94.10	4702.0
95.10	37424.0
96.10	2573.0
104.10	139.0
105.90	182.0
115.80	178.0
116.90	148.0
117.20	172.0
117.90	113.0
118.90	201.0
127.90	185.0
129.00	108.0
130.00	170.0
141.10	394.0
143.00	423.0
147.90	117.0
171.90	183.0
173.00	253.0
174.10	31984.0
175.00	2536.0
176.10	31064.0
177.00	1966.0

Instrument :
MSVOA_U
ClientSampleId :
BFB173