

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040519\
 Data File : VU030699.D
 Acq On : 04 Apr 2019 21:42
 Operator : JC/SP
 Sample : K2128-22ME
 Misc : 5.97g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF102ME

Quant Time: Apr 05 02:46:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 05 02:15:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	380012	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	361682	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	171068	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	148680	43.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.66%
7) Chloroethane-d5	1.64	69	102516	39.82	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.64%
11) 1,1-Dichloroethene-d2	2.24	63	216660	35.31	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.62%
21) 2-Butanone-d5	4.19	46	273432	101.70	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.70%
24) Chloroform-d	4.64	84	230932	44.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.12%
26) 1,2-Dichloroethane-d4	5.30	65	166566	47.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.22%
32) Benzene-d6	5.33	84	501362	47.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.96%
36) 1,2-Dichloropropane-d6	6.32	67	178300	49.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.56%
41) Toluene-d8	7.56	98	424829	45.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.78%
43) trans-1,3-Dichloropropene-	7.85	79	72844	44.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.30%
47) 2-Hexanone-d5	8.31	63	181887	109.98	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.98%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	253374	51.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.82%
64) 1,2-Dichlorobenzene-d4	11.86	152	160889	47.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.64%

Target Compounds					Ovalue
46) Tetrachloroethene	8.22	164	71707	33.976 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040519\
Data File : VU030699.D
Acq On : 04 Apr 2019 21:42
Operator : JC/SP
Sample : K2128-22ME
Misc : 5.97µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF102ME

Quant Time: Apr 05 02:46:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 05 02:15:56 2019
Response via : Initial Calibration

