

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040519\
 Data File : VU030701.D
 Acq On : 04 Apr 2019 22:28
 Operator : JC/SP
 Sample : K2246-14ME
 Misc : 6.51g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF2Q2

Quant Time: Apr 05 02:52:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 05 02:15:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	450424	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	433006	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	201243	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	173146	43.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.12%
7) Chloroethane-d5	1.64	69	118344	38.78	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.56%
11) 1,1-Dichloroethene-d2	2.23	63	254699	35.02	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.04%
21) 2-Butanone-d5	4.19	46	305770	95.95	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.95%
24) Chloroform-d	4.64	84	276660	45.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.08%
26) 1,2-Dichloroethane-d4	5.30	65	202108	48.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.44%
32) Benzene-d6	5.33	84	595534	47.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.22%
36) 1,2-Dichloropropane-d6	6.32	67	210233	48.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.06%
41) Toluene-d8	7.56	98	507726	45.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.62%
43) trans-1,3-Dichloropropene-	7.85	79	88624	44.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.72%
47) 2-Hexanone-d5	8.32	63	218216	110.21	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.21%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	302707	51.80	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	193320	48.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040519\
Data File : VU030701.D
Acq On : 04 Apr 2019 22:28
Operator : JC/SP
Sample : K2246-14ME
Misc : 6.51a/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF2Q2

Quant Time: Apr 05 02:52:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 05 02:15:56 2019
Response via : Initial Calibration

