

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040521\
 Data File : VU042928.D
 Acq On : 05 Apr 2021 14:16
 Operator : SY/MD
 Sample : M1928-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-3

Quant Time: Apr 06 01:18:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	121305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	195613	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	188670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	91162	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	97832	50.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.18%	
35) Dibromofluoromethane	5.298	113	68646	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
50) Toluene-d8	7.906	98	249640	48.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.22%	
62) 4-Bromofluorobenzene	10.639	95	90430	45.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040521\
Data File : VU042928.D
Acq On : 05 Apr 2021 14:16
Operator : SY/MD
Sample : M1928-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-3

Quant Time: Apr 06 01:18:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

