

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040619\
 Data File : VU030765.D
 Acq On : 06 Apr 2019 00:40
 Operator : JC/SP
 Sample : K2260-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETJ55

Quant Time: Apr 06 10:25:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 02:16:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	432372	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	426048	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	174034	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	209500	54.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.56%
7) Chloroethane-d5	1.68	69	159612	54.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.98%
11) 1,1-Dichloroethene-d2	2.27	63	280337	40.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.32%
21) 2-Butanone-d5	4.17	46	335411	109.65	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.65%
24) Chloroform-d	4.64	84	307053	52.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.16%
26) 1,2-Dichloroethane-d4	5.31	65	215943	53.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.36%
32) Benzene-d6	5.34	84	652814	52.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.98%
36) 1,2-Dichloropropane-d6	6.32	67	223707	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.96%
41) Toluene-d8	7.56	98	548105	49.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.44%
43) trans-1,3-Dichloropropene-	7.85	79	87641	45.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.18%
47) 2-Hexanone-d5	8.31	63	190693	97.88	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.88%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	301943	52.52	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.04%
64) 1,2-Dichlorobenzene-d4	11.86	152	183589	53.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040619\
Data File : VU030765.D
Acq On : 06 Apr 2019 00:40
Operator : JC/SP
Sample : K2260-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETJ55

Quant Time: Apr 06 10:25:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 02:16:29 2019
Response via : Initial Calibration

