

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU040619\
 Data File : VU030791.D
 Acq On : 06 Apr 2019 10:59
 Operator : JC/SP
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05096

Quant Time: Apr 08 01:09:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 01:03:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.439	0.411	6.4	101	0.00
3 T	Chloromethane	0.568	0.547	3.7	104	0.00
4 S	Vinyl Chloride-d3	0.446	0.398	10.8	93	0.00
5 T	Vinyl chloride	0.550	0.535	2.7	102	0.00
6 T	Bromomethane	0.285	0.276	3.2	102	0.00
7 S	Chloroethane-d5	0.339	0.328	3.2	101	0.00
8 T	Chloroethane	0.314	0.322	-2.5	110	0.00
9 T	Trichlorofluoromethane	0.641	0.638	0.5	106	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.358	0.329	8.1	100	0.00
11 S	1,1-Dichloroethene-d2	0.807	0.775	4.0	101	0.00
12 T	1,1-Dichloroethene	0.346	0.344	0.6	104	0.00
13 T	Acetone	0.362	0.282	22.1	82	0.00
14 T	Carbon disulfide	1.074	1.037	3.4	102	0.00
15 T	Methyl Acetate	0.561	0.592	-5.5	109	0.00
16 T	Methylene chloride	0.420	0.414	1.4	104	0.00
17 T	trans-1,2-Dichloroethene	0.367	0.350	4.6	101	0.00
18 T	Methyl tert-butyl Ether	1.229	1.223	0.5	105	0.00
19 T	1,1-Dichloroethane	0.757	0.780	-3.0	108	0.00
20 T	cis-1,2-Dichloroethene	0.405	0.410	-1.2	107	0.00
21 S	2-Butanone-d5	0.354	0.388	-9.6	110	0.00
22 T	2-Butanone	0.424	0.435	-2.6	100	0.00
23 T	Bromochloromethane	0.192	0.197	-2.6	107	0.00
24 S	Chloroform-d	0.682	0.693	-1.6	105	0.00
25 T	Chloroform	0.761	0.767	-0.8	107	0.00
26 S	1,2-Dichloroethane-d4	0.465	0.477	-2.6	106	0.00
27 T	1,2-Dichloroethane	0.581	0.601	-3.4	110	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
29 T	Cyclohexane	0.713	0.695	2.5	99	0.00
30 T	1,1,1-Trichloroethane	0.609	0.634	-4.1	107	0.00
31 T	Carbon tetrachloride	0.525	0.546	-4.0	107	0.00
32 S	Benzene-d6	1.460	1.453	0.5	101	0.00
33 T	Benzene	1.636	1.711	-4.6	107	0.00
34 T	Trichloroethene	0.401	0.413	-3.0	107	0.00
35 T	Methylcyclohexane	0.619	0.586	5.3	97	0.00
36 S	1,2-Dichloropropane-d6	0.500	0.506	-1.2	102	0.00
37 T	1,2-Dichloropropane	0.463	0.485	-4.8	108	0.00
38 T	Bromodichloromethane	0.559	0.587	-5.0	109	0.00
39 T	cis-1,3-Dichloropropene	0.659	0.667	-1.2	100	0.00
40 T	4-Methyl-2-pentanone	0.745	0.818	-9.8	108	0.00
41 S	Toluene-d8	1.294	1.320	-2.0	100	0.00
42 T	Toluene	1.670	1.764	-5.6	106	0.00
43 S	trans-1,3-Dichloropropene-d	0.228	0.227	0.4	98	0.00
44 T	trans-1,3-Dichloropropene	0.581	0.601	-3.4	103	0.00
45 T	1,1,2-Trichloroethane	0.393	0.413	-5.1	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU040619\
 Data File : VU030791.D
 Acq On : 06 Apr 2019 10:59
 Operator : JC/SP
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05096

Quant Time: Apr 08 01:09:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 01:03:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Tetrachloroethene	0.292	0.281	3.8	101	0.00
47 S	2-Hexanone-d5	0.229	0.265	-15.7	109	0.00
48 T	2-Hexanone	0.581	0.654	-12.6	106	0.00
49 T	Dibromochloromethane	0.417	0.440	-5.5	109	0.00
50 T	1,2-Dibromoethane	0.425	0.438	-3.1	106	0.00
51 T	Chlorobenzene	1.024	1.079	-5.4	109	0.00
52 T	Ethylbenzene	1.821	1.890	-3.8	106	0.00
53 T	m,p-Xylene	0.635	0.657	-3.5	104	0.00
54 T	o-xylene	0.632	0.661	-4.6	105	0.00
55 T	Styrene	1.083	1.172	-8.2	105	0.00
56 T	Isopropylbenzene	1.663	1.749	-5.2	105	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.675	0.721	-6.8	106	0.00
58 T	1,1,2,2-Tetrachloroethane	0.711	0.733	-3.1	105	0.00
59	1,2,3-Trichloropropane	0.555	0.589	-6.1	106	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
61 T	Bromoform	0.661	0.679	-2.7	109	0.00
62 T	1,3-Dichlorobenzene	1.580	1.568	0.8	105	0.00
63 T	1,4-Dichlorobenzene	1.607	1.633	-1.6	109	0.00
64 S	1,2-Dichlorobenzene-d4	0.983	0.986	-0.3	104	0.00
65 T	1,2-Dichlorobenzene	1.585	1.651	-4.2	109	0.00
66 T	1,2-Dibromo-3-chloropropane	0.337	0.353	-4.7	108	0.00
67	1,3,5-Trichlorobenzene	1.155	1.113	3.6	100	0.00
68 T	1,2,4-trichlorobenzene	0.982	1.005	-2.3	103	0.00
69	Naphthalene	3.183	3.362	-5.6	103	0.00
70 T	1,2,3-Trichlorobenzene	1.026	1.053	-2.6	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0